



8-Pin Mini-DIN Standard Pinout

Always identify proper wiring via continuity check & color when using multi-strand cables.

The calibration certificate for the device should be used as the definitive reference for custom wiring options.



Female Connector (Device)



Male Connector (Cable)

PIN	8-PIN MINI-DIN CONTROLLER
1	NC: Not configured for use, do not connect. Optional: 4–20 mA analog output signal.
2	Analog Out 2: Static 5.12 Vdc. Optional: Analog signal to indicate another parameter (0–5 Vdc, 1–5 Vdc, 0–10 Vdc, or 4–20 mA)
3	Rx or B (+): Receives RS-232 (Rx) or RS-485 B (+) signals to change the device's settings.
4	Analog In: 0–5 Vdc analog DC input defining the setpoint. Optional: 1–5 Vdc, 0–10 Vdc, or 4–20 mA
5	Tx or A (-): Sends RS-232 (Tx) or RS-485 A (-) signals from the device.
6	Analog Out: 0–5 Vdc output signal. Optional: 1–5 Vdc, 0–10 Vdc
7	Power In: Powers the device, see the specification sheet for details.
8	Ground: Common ground for power, digital communications, analog signals and alarms.

PIN	8-PIN MINI-DIN METER
1	Not Used: Not configured for use, do not connect. Optional: 4–20 mA analog output signal.
2	Analog Out 2: Static 5.12 Vdc. Optional: Analog signal to indicate another parameter (0–5 Vdc, 1–5 Vdc, 0–10 Vdc, or 4–20 mA)
3	Rx or B (+): Receives RS-232 (Rx) or RS-485 B (+) signals to change the device's settings.
4	Ground to Tare: Ground this pin to tare the device.
5	Tx or A (-): Sends RS-232 (Tx) or RS-485 A (-) signals from the device.
6	Analog Out: 0–5 Vdc output signal. Optional: 1–5 Vdc, 0–10 Vdc
7	Power In: Powers the device, see the specification sheet for details.
8	Ground: Common ground for power, digital communications, analog signals and alarms.

Note: Do not connect RS-485 to RS-232 units or cables. Damage will occur. Check part number or contact factory to verify RS-485 functionality.